

SAMPLING INSTRUCTIONS – Lead Analysis

Container Type: 1 liter (32 oz) plastic or glass

1. On the chain of custody, please fill out your name, address, phone number, and answer any questions concerning the water source prior to submitting your sample to the laboratory for analysis. Due to confidentiality requirements, Lozier Environmental Consulting, Inc. can only release test results to the person submitting the sample. Otherwise, special instructions must be indicated on the form.
2. Allow water to stand in the pipes for 6 – 8 hours (preferably overnight).
3. Remove all filters and screens from the faucet and/ or water source. Open the top of the bottle and take the first sample of water from the source (first draw).
4. Fill the bottle. Make sure the lid is on tight and refrigerate the sample.
5. Submit the sample within **24 HOURS** of collection.
6. Samples **MUST BE DELIVERED TO THE LAB** no later than 4:00 pm Monday through Thursday.
7. The results will be mailed within 14 business days from receipt of the sample at Lozier Environmental Consulting, Inc.
8. The laboratory is closed weekends and all national holidays and extended holiday weekends.

If you have any questions regarding these sampling instructions, please call Lozier Environmental Consulting, Inc. at (585) 654-9080.

