

SAMPLING INSTRUCTIONS – Nitrate & Nitrite

Container Type: 250 ml (8 oz) plastic or glass

1. On the chain of custody, please fill out your name, address, phone number, and answer any questions concerning the water source prior to submitting your sample to the laboratory for analysis. Due to confidentiality requirements, Lozier Environmental Consulting, Inc. can only release test results to the person submitting the sample. Otherwise, special instructions must be indicated on the form.
2. Remove all filters and screens from the faucet and/ or water source.
3. Fill the bottle. Make sure the lid is on tight and refrigerate the sample.
4. Transport the sample in a cooler with ice, keeping the sample under 20° C. Submit the sample within **24 HOURS** of collection.
5. Samples **MUST BE DELIVERED TO THE LAB** no later than 4:00 pm Monday through Thursday.
6. The results will be mailed within 10 - 15 business days from receipt of the sample at Lozier Environmental Consulting, Inc.
7. The laboratory is closed all national holidays and extended holiday weekends.

If you have any questions regarding these sampling instructions, please call Lozier Environmental Consulting, Inc. at (585) 654-9080.